बिड संख्या/Bid Number: GEM/2026/B/8158545
दिनांक /Dated: 16-10-2026

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	02-11-2026 14:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	03-11-2026 14:00:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Chhattisgarh
विभाग का नाम/Department Name	Health And Family Welfare Department Chhattisgarh
संगठन का नाम/Organisation Name	Chhattisgarh Medical Services Corporation Limited
कार्यालय का नाम/Office Name	Chhattisgarh Medical Services Corporation Limited
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :md.cgmsc@gov.in Buyer Email id: bme.cgmsc2001@gmail.com
कुल मात्रा/Total Quantity	519
वस्तु श्रेणी /Item Category	Microscopes - Pathological and Research as per IS 4381, IS 5204, IS 4381, IS 5204 (Q3)
पिछले 3 वित्तीय वर्षों का बिडर का न्यूनतम औसत वार्षिक वित्तीय टर्नओवर/Minimum Average Annual Turnover of the bidder (For the last 3 Financial Years)	90 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	180 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
टर्नओवर पात्रता मानदंड/Turnover eligibility criteria	To be verified by the buyer at the time of technical evaluation
वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover	Yes | Complete
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover	Yes | Complete

बिड विवरण/Bid Details	
विक्रेता से मांगे गए दस्तावेज़/**Document required from seller**	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/**Do you want to show documents uploaded by bidders to all bidders participated in bid?**	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / **Minimum number of bids required to disable automatic bid extension**	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / **Number of days for which Bid would be auto-extended**	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / **Number of Auto Extension count**	1
विगत प्रदर्शन /**Past Performance**	40 %
बिड से रिवर्स नीलामी सक्रिय किया/**Bid to RA enabled**	No
बिड का प्रकार/**Type of Bid**	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/**Primary product category**	Microscopes - Pathological and Research as per IS 4381, IS 5204, IS 4381, IS 5204
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /**Time allowed for Technical Clarifications during technical evaluation**	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/**Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)**	No
Payment Timelines	Payments shall be made to the Seller within **45** days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/**Evaluation Method**	Total value wise evaluation
मध्यस्थता खंड/**Arbitration Clause**	No
सुलह खंड/**Mediation Clause**	No

ईएमडी विवरण/EMD Detail

एडवाईजरी बैंक/Advisory Bank	HDFC Bank
ईएमडी राशि/EMD Amount	182590

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	42

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :

MD
Chhattisgarh Medical Services Corporation Limited, Health and Family Welfare Department Chhattisgarh,
Chhattisgarh Medical Services Corporation Limited,
(Managing Director)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% / Purchase Preference to MSE OEMs/ Service Provider available upto price within L1+X%	15

सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
5. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
7. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
8. **Purchase preference to Micro and Small Enterprises (MSEs):** Purchase preference will be given to MSEs from the State of Bid Inviting Authority having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry/ State of Bid Inviting Authority. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry. Benefits of MSE

will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
9. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
10. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 40% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
21-10-2026 15:00:00	CGMSC OFFICE Sadar Vallabh Bhai Patel Bhawan Kayabandha Nava Raipur Sec 24

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (325 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	325	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Jagdish Prasad Singh	497001,Office of the Dean,Rajmata Shrimati Devendra Kumari Singhdeo Govt. Medical College, Kanya Parisar Road, Gangapur,Ambikapur, Surguja	1	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (92 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Deepti Gautam	493445,Govt. Medical College Villege- Kharora Mahasamund	92	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (30 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	30	55

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Jagdish Prasad Singh	497001,Office of the Dean,Rajmata Shrimati Devendra Kumari Singhdeo Govt. Medical College, Kanya Parisar Road, Gangapur,Ambikapur, Surguja	1	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (12 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Suresh Kumar Thakur	494001,Office of the Dean, Late Baliram Kashyap Memorial Govt. Medical College, Dimrapal, Jagdalpur	12	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (50 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display
	Eye piece Type	Huygenian, Compensating, NA, Ramsden

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Suresh Kumar Thakur	494001,Office of the Dean, Late Baliram Kashyap Memorial Govt. Medical College, Dimrapal, Jagdalpur	50	60

Microscopes - Pathological And Research As Per IS 4381, IS 5204, IS 4381, IS 5204 (4 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification]

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	**Microscope Type**	pathological monocular hinged type without built in light and without light intensity regulator, pathological monocular non hinged type with built in light and with light intensity regulator, pathological binocular non hinged type with built in light and with light intensity regulator, research binocular non hinged type with built in light and with light intensity regulator, Research Trinocular - Non-Hinged type - with built-in light and with light intensity regulator, Compact integrated unit including: cell imaging system, digital camera, computer, high power fluorescence lighting system and LCD display, research/ pathology with display

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Eye piece Type	Huygenian, Compensating, NA, Ramsden
	Eye Piece with magnification	monocular 5 x, monocular 10 x, monocular 15 x, set of two for binocular 5 x, Set of Two For Binocular 10x, Internal magnification 23x, NA
	Objective Type	achromatic, semiapochromatic, Apochromatic, Semi Plan Achromatic, Plan Achromatic, Plan Achromatic Infinity
	Objective Magnification	10 x, 20 x, 40 x, 45 x, 90 x, 100X, 4X
	Numerical Aperture of Objective	0.25, 0.5, 0.65, 1.25, 0.45, 0.10
ACCESSORIES	**Extra eyepieces (inclusive in the scope of supply)**	5 x, 10 x, 15 x, set of two 5 x, set of two 10 x, set of two 15 x, NA
WARRANTY	**Warranty**	1, 2, 3, 4, 5 Or higher **(year)**

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Kuldip Singh Sangha	490024,Chandulal Chandrakar Memorial Government Medical College Kurud Road, Kachandur, Dist. Durg, Chhattisgarh - 490024, India	4	60

Special terms and conditions-Version:3 effective from 14-01-2025 for category Microscopes - Pathological and Research as per IS 4381, IS 5204, IS 4381, IS 5204

1. **Special Terms and Conditions**

    1. **Comprehensive warranty**

    Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

    2. **Service centres** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online

registration of service requests also to be provided buyer/consignee at the time of supplies

3. **Source of supply**: It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them .
4. **Packing and Marking** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details
5. **Spare Parts** : Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied

6. **Installation, Training, Manuals** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable
7. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
8. Software: All software updates should be provided free of cost during warranty period

Note- The product may also carry an ISI Mark/Standard Mark as per the BIS standard governing this category. As such Buyers may also opt for ISI marked product while creating BID. In such cases buyer shall verify valid BIS License for the same at the time of evaluation of BID and check ISI marking at the time of receipt of material before issuing CRAC.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. **Generic**

   OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document Click here to view the file.

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid.All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents (other than EMD and signed Integrity Pact - which can be submitted within 5 days of bid opening) in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---